

Material Certificate

Wir bestätigen hiermit, dass die
We hereby certify that the

ultrafilter Medizinal-Vakuumfilterelemente
ultrafilter medical vacuum filter elements
FFMV



Filterelemente, ungeachtet der Ausführung sind mit Qualitäts-Komponenten hergestellt, welche nicht giftig und biologisch inerte Rohmaterialien sind und den Vorschriften der FDA gemäß CFR, title 21, EC/1935/2004 entsprechen.

are manufactured from high-quality, non-toxic and biologically inert raw materials and components and are FDA listed for food contact as per Code of Federal Regulation (CFR), Title 21, EC/1935/2004 as listed below:

Filtermedium Filter media	Borosilikatglassfieber Borosilicate glass fibers	CFR title 21, 211.72 CFR title 21, 177.2260
Endkappen End caps	Edelstahl Stainless steel	CFR title 21, 211.65
Dichtringe O-rings	Viton	CFR title 21, 177.2600
Vergussmasse Sealing compound	Polyurethan polyurethane	CFR title 21, 177.1680
Stützmantel Support core	Edelstahl Stainless steel	CFR title 21, 211.65



Hilden, May 2017